



www.venturemedical.com

Venture Medical ReQuip, Inc.
6008 Bonacker Drive
Tampa, FL 33610
800-627-3215 | 813-740-8640
Fax: 800-466-1251
sales@venturemedical.com

Biolight M9000A Color Patient Monitor - New



The Biolight M9000A Patient Monitor features a 12.1" color LCD display with user friendly menus and a navigator knob for easy configurations.

The M9000A is a perfect solution for recovery, PACU, procedure rooms and anywhere a compact multi-parameter monitor is needed.

Standard configuration includes: ECG, NIBP, Resp, SpO2, PR, 2-Temp, Co2 Ready, Printer, and VGA Output.

This multi-parameter monitor has ST arrhythmia analysis, anti high-frequency interference, anti-defibrillation, isolated float and networking capacity.

The Biolight M9000A includes: ECG cable and leadwires (3 lead), NIBP cuff, NIBP hose, SpO2 finger sensor, lithium battery, recorder paper, and user manual.

Optional rolling stands and wall mounts are also available.

Manufacturer Warranty: 2 Years, 90 Days for Accessories

Technical Specifications:

- Display: 12.1" Color LCD
- Dimension: 12.5" (W) x 10.4" (H) x 6" (D)
- Weight: ≤ 9.9 lbs. (with battery)
- Power Supply: 120~240V; AC: 50~ 60 Hz
- Alarm Indication: Visual and Audible Alarm

Trend

- Store & Review: 168 hours trend data and trend maps
- Parameter option: SpO2, NIBP, PR, Resp, Co2, Temp1-2, IBP1-2, ST.
- Storage Cycle Intervals: 1-5min, 10min, 15min, 20min, 25min, 30min.

Parameter Specifications:

ECG

- Leads: I, II, III, aVR, aVL, aVF
- Mode: 3-leads, 5-leads
- Gains: 2.5mm/mV, 5mm/mV, 10mm/mV, 20mm/mV, 40mm/mV, auto
- Scanning Speed: 12.5mm/s, 25mm/s, 50mm/s
- Heart Rate Range:
 - Adult: 10~300bpm
 - Child & Neonate: 10~350bpm
- Accuracy: ±1bpm or ±1% whichever is greater

ST

- Measurement Range -2.0mV – 2.0mV
- Resolution 0.01mV

NIBP

- Operation Modes: Automatic Oscillometry
- Range: Adult: 10 ~270mmHg
Child: 10 ~235mmHg
Neonate: 10 ~135mmHg
- Measurement: Systolic Pressure/Diastolic Pressure/Mean Pressure
- Unit: mmHg/Kpa
- Pulse Rate Range: 40 ~240bpm

RESP

- Method: Impedance variation between RA-LL (R-F)
- Measuring Range: 0~150rpm
- Accuracy: ±2 rpm
- Gain: x1, x2, x4
- Sweep speed 6.25mm/s, 12.5mm/s, 25mm/s

Biolight SpO2 (Pulse Oximetry)

- Range: 0 ~100%
- Accuracy: ±2% (70% ~100%); Unspecified (0 ~69%)
- Range & Accuracy of Pulse Rate: 25 ~250bpm ±1bpm or ±1% whichever is greater

TEMP

- Channel: 2 channel
- Range: 32~122°F, 0~50°C
- Unit: Celsius, Fahrenheit

Standard Configuration:

- ECG, RESP, NIBP, SpO2, PR, Temp, VGA Output, Co2 Ready, Rechargeable Lithium Battery, Printer

IBP (option)

- Channel: 2
- Measurement Range: -50 ~+300mmHg
- Accuracy: ±2mmHg or 2%, whichever is greater

EtCo2 (option, Sidestream)

- Range 0 ~19.7% (0 ~150mmHg)
- Unit: %,mmHg, kPa
- Respiration Rate Range 2 ~150bpm

VENTUREMEDICAL.COM

© 2007-2025 Venture Medical ReQuip, Inc.